



“औद्योगिक, पर्यटकीय र हरित नगर, स्वच्छ, सुरक्षित र समृद्ध नगर”

दमक नगरपालिका नगर कार्यपालिकाको कार्यालय



प.सं.:

च.नं.:

दमक, भद्रा
कोशी प्रदेश, नेपाल



☎: ०२३-५९७०००

मिति: २०८१/१२/१९

श्री सम्पूर्ण फर्महरू,

विषय: दर रेट पेश गर्ने सम्बन्धमा।

प्रस्तुत विषयमा दमक नगरपालिकाद्वारा स्थापना गरिएको खाद्य परिक्षण प्रयोगशाला संचालन गर्न केहि उपकरण, रसायन तथा अन्य सामग्रीहरूको लागत अनुमान तयार पार्न आवश्यक भएको हुँदा संलग्न सुचि अनुसारको सामान तथा कार्यको दररेट मिति २०८१/१२/२४ गते १२ बजे भित्र दमक नगरपालिकामा पेश गर्नु हुन अनुरोध छ।

आवश्यक कागजातहरू:

फर्म दर्ता प्रमाणपत्रको प्रतिलिपि
मुअकर दर्ता प्रमाणपत्रको प्रतिलिपि
अघिल्लो आ.व.को कर चुक्ता
दररेट समावेस भएको कागजात

१२/१२/१९
यशराज निरौला
वरिष्ठ प्रशासकीय अधिकृत

Price Schedule of equipment's

S.No	Name	Work Description of Equipment	Unit	Quantity	Rate	Rate in Words	Total
1	Lactometer	For determining specific gravity of milk ranging from 1.020-1.035 at 27°C and a surface tension of 50 mN/m.	pic	3			
2	Fume hood	Internal size (W*D*H): 1020*670*730 with air velocity: 0.3-0.8m/s, noise: ≤70 dB, chemical resistance working table, UV lamp installed, with LED display and waterproof socket.	set	1			
3	Milk adulteration test kit	Manual testing large size kit detecting components as Urea, Starch, Salt, Ammonia, Sugar, Neutralizers, Nitrate, Glucose and Hydrogen Peroxide.	set	1			
4	Thermo meter	Portable and digital thermometer with temperature range: -10 to 150 °C, temperature accuracy of ± 0.1°C and length of probe ≥125 mm	pic	1			
Total							
VAT 13%							
Total with VAT							

Price Schedule of chemicals and other devices

S.No.	Name	Unit	Quantity	Rate	Rate in Words	Total
1	Distill water (with zero conductivity)	bottle	1			
2	Gerber sulphuric acid	bottle	1			
3	Iso-amyl alcohol	bottle	1			
4	Sodium thiosulphate	bottle	1			
5	Potassium dichromate	bottle	1			
Total						
VAT 13%						
Total with VAT						

Price Schedule of other devices

S.No.	Name	Work Description of Equipment	Unit	Quantity	Rate	Rate in Words	Total
1	Spatula	Non-corrosive material, having both spoon and flat ends, length: ≤ 20 cm.	pic	3			
2	Glass rod	Solid glass rod, length: ≤ 20 cm	pic	3			
3	Aluminum tray	Tray size (l*b*h): (≤ 30 * ≤ 21 * ≤ 3 cm)	pic	2			
4	Microwave gloves	Heat resistance	pair	1			
5	Safety goggles	Used in chemistry lab	pair	1			
6	Laboratory eyewash	Control flow of water to both eye simultaneously with a minimum flow rate of (1.5 liters/minute) for 10 min.	set	1			
Total							
VAT 13%							
Total with VAT							

.....
Signature

Name of Firm

Seal of Firm