



“औद्योगिक, पर्यटकीय र हरित नगर, स्वच्छ, सुरक्षित र समृद्ध नगर”

दमक नगरपालिका

नगर कार्यपालिकाको कार्यालय



प.सं.: ०८२/०८३
च.नं.: ४६९२



☎: ०२३-५९७०००

श्री आपूर्तिकर्ताहरू सबै

मिति २०८२।१२।०४

विषय : मादक पदार्थ सेवन जांच गर्ने बेथलाइजर मेशिन तथा लागु पदार्थ सेवन जांच गर्ने कीटको प्रति इकाई लागत मुल्य (दररेट) पेश गर्ने सम्बन्धि संशोधित सूचना ।

प्रस्तुत विषयमा ईलाका प्रहरी कार्यालय दमकको लागि मादक पदार्थ सेवन जांच गर्ने बेथलाइजर मेशिन तथा लागु पदार्थ सेवन जांच गर्ने कीटको स्वरिद प्रकृया सम्पादन गर्नुपर्ने भएको हुदा उक्त दुवै सामाग्रीको प्राबिधिक स्पेशिफिकेशन सूचना साथ संलग्न रहेको छ ।

सप्लाई गर्न ईच्छुक आपूर्तिकर्ताहरूले स्पेशिफिकेशन बमोजिमका सामाग्रीको प्रति इकाई लागतमुल्य (दररेट) मिति २०८२।१२।०९ गते भित्रमा पेश गर्न यो संशोधित सूचना प्रकाशित गरिएको छ ।

यस कार्यालयमा सूचीकृत नभएका फर्म वा आपूर्तिकर्ताहरूले उल्लेखित समय सिमा भित्र नियमानुसार सूचीकृत गरी दररेट पेश गर्न सकिने समेत सम्बन्धित सबैको जानकारीका लागि यो सूचना प्रकाशित गरिएको छ ।

नोट : बेथलाइजर तथा कीटको स्पेशिफिकेशन सूचना साथ संलग्न रहेको छ ।

बोधार्थ :

श्री सूचना तथा प्रबिधि शाखा (नगरपालिकाको वेवसाइट तथा फेसबुक पेजमा अपलोड गर्नु हुन ।)

सुनिल निरौला
प्रमुख प्रशासकीय अधिकृत




सुनिल निरौला
 अधिकृत

Technical Specification of Breath Alcohol Analyzer

S.NO	ITEMS	DEPARTMENTAL REQUIREMENT	Remarks (Brand, Model No, Country of Origin and Compliance)
1	Breath Alcohol Analyzer	<ul style="list-style-type: none"> • Type: Alcohol Analyzer • Sensor: Electrochemical fuelcell/infrared (IR) • Detection range: (0-0.4%BAC)/(0-2.0mg/L BrAC) or Better • Operation: Android/ Linux • Memory: min 50,000 tests records storage in memory • Blowing Time: less than or equal to 10 second for detailed operation and less than or equal to 3 sec for rapid operation • Test response time: less than or equal to 6 second for rapid operation and less than or equal to 30 seconds for detailed • Display screen: minimum 3.5 inch colorful touch screen • Communication: 4G, wifi • Satellite navigation system: GPS/ Glonass/ Baidou or equivalent • Printer: built-in micro thermal printer • Battery: Easily detachable and replaceable 2000mAh • Camera : Built-in camera for taking clear picture of the suspect in day/night • Test : Normal Test mouthpiece-200pcs • Rapid Screening Mouth piece-10 pcs • Thermal Paper roll-20 rolls • Power charger -1 pc • Charging dock - 1 pc • Extra battery-1 pc • Car mounted charger – 1 pc • Pouch -1 pc • USB cable- 1 pc , able to transfer the data and be printed when required • Instrument box-1 pc • Warranty: 2 Year, free Calibration Service for at least 2 years, • Certification and Authorization: CE, ISO, Manufacturer/Dealer Authorization. • Catalogue must be submitted and serial number and model number must be labelled on hardware. • Bidder must provide all necessary software, orientation and training to operate the device. 	
<p>Note: The major item is indicated by bold letters/alphabets/figures in the technical specifications. Non-compliance of these major items will result in non-responsiveness of the bid offered.</p>			

Per Unit Rate:

Per Unit Rate in Words:



Technical Specification of Drug Test KIT

सुनिल निरौला
प्रमुख प्रशासकीय अधिकृत

S.N.	Description	Requirement	Per Unit Price Without Vat
1.	cut-off	THC 50 ng/ml AMP 1000 ng/ml BAR 300 ng/ml BZO 300 ng/ml COC 300 ng/ml OPI/MOP 300 ng/ml MTD 300 ng/ml MAMP/MET 1000 ng/ml MDMA 500 ng/ml BUP 10 ng/ml OXY 100 ng/ml OTHERS (eg. PCP etc) varies 99% accurate when used as directed	
2.	Materials Required	> Test Kit > Specimen [Urine] collection container > Disposable dropper	
3.	Stored Temperature	Minimum [5-28 degree C }	
4.	Features	> Expiration date on test kit must be label on the sealed pouch. > The test kit must remain in the sealed pouch until use.	

NOTE: MUST have at least 8 parameters